

- HCT is directly measured.
- As the RBCs pass through the aperture, the volume of individual RBCs are measured.
- These measurements are added up and compared to the volume of the whole blood.
- This produces a HCT that closely matches the reference method, or spun HCT result

